

Turbi Check with infra-red light source (EN ISO 7027)



The compact Lovibond® infrared turbidity meter TurbiCheck is designed to allow fast, precise onsite testing. The unit measures the scattered light at an angle of 90°, as stipulated in EN ISO 7027.

The wide measuring range from 0.01-1100 TE/F = NTU = FNU makes the instrument suitable for various applications, ranging from drinking water to waste water.

As infrared light is used for measurement purposes, the unit can be used to test both coloured and colourless liquids.

The standards required for calibration of the unit are also supplied. A second adjustment mode allows alternative adjustment with user-defined turbidity standards.

Accessories

Turbidity standard set T-CAL (<0.1, 20, 200, 800 NTU)

Order code: 19 41 50

Set of 12 empty sample vials, 24 mm ø

Order code: 19 76 55

Highlights

- Range 0.01 1100 NTU
- Measurement with infrared light at an angle of 90°
- Measurement of coloured liquids
- Easy handling
- 600 tests without battery change

Delivery Content

- Instrument in carrying case
- 4 turbidity standards (<0,1, 20, 200 and 800 NTU)
- 9 V battery

- 3 vials (ø 24 mm) with lida
- Guarantee sheet
- Certificate of Compliance
- Instruction Manual

Order code: 26 60 20

Technical data

Measurement cycle	approx. 8 seconds
Display	backlit LCD (on keypress)
Optics	temperature- compensated LED (λ = 860 nm) and photosensor amplifier in water proof sample chamber, infrared light
Keypad	polycarbonate membrane, splash proof
Power supply	9 V power pack battery
Auto - OFF	automatic switch-off
Storage	internal ring memory for 16 data sets
Additional feature	real time clock and date
Range (Auto-range)	0,01 - 1100 NTU
Resolution	0.01 - 9.99 NTU = 0.01 NTU 10.0 - 99.9 NTU = 0.1 NTU 100 - 1100 NTU = 1 NTU
Accuracy	± 2,5 % of reading or ± 0.01 NTU (0 - 500 NTU) ± 5 % (500 - 1100 NTU)
Housing	ABS
Dimensions (L x W x H)	190 x 110 x 55 mm
Weight (base unit)	approx. 0.4 kg
Ambient conditions	Temperature: 5 – 40 °C rel. humidity: 30 – 90%

CE-Conformity

General Catalogue March 2010