# MD 100 Photometer



# Small I Mobile I Rapid

The MD 100 uses high quality interference filters with long-life LEDs as a light source without any moving parts in a transparent sample chamber.

The units supply accurate, reproducible results very quickly. Other major advantages include ease of operation, ergonomic design, compact dimensions and safe handling.

The calibration and software-based adjustment options mean that the MD 100 is also suitable for use as a testing instrument for other equipment

The tests are conducted using either Lovibond® tablet reagents with long-term stability and a guaranteed minimum 5 or 10 year shelf life, VARIO powder reagents or liquid reagents.

0

Please see pages 88 onwards for reagents (order codes)

## Highlights

- Drift-free results through high quality interference filter ensured
- Scroll memory
- Automatic switch-off
- Real-Time-Clock and date
- Calibration mode
- Backlit display
- Storage function
- One Time Zero (OTZ)
- Waterproof\*)

\*) as defined in IP 68, 1 hour at 0.1 meter

Single-Parameter		Single-Parameter		4in1		
Test	Code	Test	Code	Test	Code	
MD 100 Aluminium, tablet reagents 0.01 - 0.3 mg/l Al	27 62 00	<b>MD 100 Molybdenum LR</b> Powder reagents / reagent solution	27 61 40	MD 100 Chlorine, pH, Cyanuric acid, Alkalinity-M,	27 80 70	
<b>MD 100 Aluminium</b> , powder reagents	27 62 05	0.03 - 3.0 mg/l Mo (mixing cylinder renot included)		0.02 - 6.0 mg/l Cl <sub>2</sub> / 0,1 - 10 mg/l	tablet reagents <b>(OTZ)</b> 0.02 - 6.0 mg/l Cl <sub>2</sub> / 0,1 - 10 mg/l Cl <sub>2</sub> *	
0.01 - 0.25 mg/l Al		MD 100 Molybdenum HR, powder reagents	27 61 41	6.5 - 8.4 pH ; 0 - 160 mg/l cyanuri 5 - 200 mg/l CaCO₃ (TA)	c acid	
<b>MD 100 Ammonia</b> , tablet reagents 0.02 - 1.0 mg/l N	27 60 60	0.3 - 40 mg/l Mo	27.61.12	MD 100 Chlorine, pH, 27 80 75 Cyanuric acid, Alkalinity-M (total) liquid reagent for chlorine and pH (OTZ) 0.02 - 4 mg/l Cl <sub>2</sub> / 6.5 - 8.4 pH 0 - 160 mg/l cyanuric acid / 5 - 200 mg/l CaCO <sub>3</sub> (TA)		
<b>MD 100 Ammonium</b> , powder reagents 0.01 - 0.8 mg/l N	27 60 65	MD 100 Molybdenum, tablet reagents 0.6 - 30 mg/l Mo	27 61 42			
<b>MD 100 Chloride</b> , tablet reagents 0.5 - 25 mg/l CΓ	27 61 80	<b>MD 100 Phosphate</b> , tablet reagents 0.05 - 4.0 mg/l PO <sub>4</sub>	27 60 40	MD 100 Chlorine Duo, Chlorine HR, 27 81 60 pH, Alkalinity-M, Calcium hardness powder reagents and tablet reagents for Chlorine tablet reagents for pH, Alkalinity-M, Calcium hardness without reagents for Chlorine HR Chlorine T 0,01 - 6,0 mg/l Chlorine PP 0,02 - 3,5 mg/l Chlorine HR 5 - 200 mg/l pH 6,5 - 8,4 Alkalinity-M 5 - 200 mg/l Calcium hardness 20 - 500 mg/l		
5 - 250 mg/l Cl <sup>-</sup> (by dilution) <b>MD 100 Chlorine</b> ,	27 60 00	MD 100 Phosphate, powder reagents	27 60 45			
tablet reagents <b>(OTZ)</b> 0.01 - 6.0 mg/l Cl <sub>2</sub> / 0.1 - 10 mg/l Cl <sub>2</sub> *		0.06 - 2.5 mg/l PO <sub>4</sub> <b>MD 100 Silica</b> , tablet reagents	27 61 10			
<b>MD 100 Chlorine</b> , liquid reagent <b>(OTZ)</b> 0.02 - 4 mg/l Cl <sub>2</sub>	27 60 05	0.05 - 4.0 mg/l SiO <sub>2</sub> <b>MD 100 Silica LR</b> , powder reagents 0.1 - 1.6 mg/l SiO <sub>2</sub>	27 61 15			
MD 100 Chlorine DUO, for 2 types of 1) Tablet reagents	f reagents 27 60 20	MD 100 Silica HR, powder reagents 1 - 90 mg/l SiO <sub>2</sub>	27 61 16			
0.01 - 6.0 mg/l Cl₂ / 0,1 - 10 mg/l Cl 2) Powder reagents 0.02 - 2.0 mg/l Cl₂ (ø 24 mm glass v	27 60 25 ia <b>l</b> )	MD 100 Suspended solids no reagents required 0 - 750 mg/l TSS	27 61 50			
0.1 - 8.0 mg/l Cl <sub>2</sub> (ø 10 mm <b>multi vial-2</b> ) <b>MD 100 Chlorine</b> , powder reagents 27 60 10  0.02 - 2.0 mg/l Cl <sub>2</sub> (ø 24 mm glass vial)		<b>MD 100 Urea</b> , tablet reagents 0.1 - 2.5 mg/l Urea 0.2 - 5 mg/l Urea (by dilution)	27 62 10	<b>5in1 MD 100 Chlorine, pH,</b> 27 80 80		
0.1 - 8.0 mg/l Cl <sub>2</sub> (ø 10 mm <b>multi vial-2</b> ) <b>MD 100 Chlorine HR</b> 27 61 70 <b>(Potassium iodide),</b> tablet reagents 5 - 200 mg/l Cl <sub>2</sub> (ø 16 mm round vial & adapter)				Cyanuric acid, Alkalinity-M, Calcium hardness tablet reagents (OTZ) $0.02 - 6.0 \text{ mg/l Cl}_2 / 0,1 - 10 \text{ mg/l Cl}_2*$ 6.5 - 8.4  pH; $0 - 160  mg/l cyanuric acid$		
MD 100 Chlorine dioxide,	27 60 30	2in1		$5 - 200 \text{ mg/l CaCO}_3 (TA)$ ; $0 - 500 \text{ ng}$		
tablet reagents 0.02 - 11 mg/l ClO <sub>2</sub>		MD 100 Chlorine, pH,	27 80 20			
MD 100 Chlorine dioxide, powder reagents	27 60 35	tablet reagents <b>(OTZ)</b> 0.01 - 6.0 mg/l Cl <sub>2</sub> / 0,1 - 10 mg/l Cl <sub>2</sub> * 6.5 - 8.4 pH		6in1		
0.04 - 3.8 mg/l CIO <sub>2</sub> MD 100 COD, tube tests, without read 27 61 20	gents	MD 100 Chlorine, pH, liquid reagent (OTZ)	27 80 25	MD 100 Chlorine, Bromine, pH, Cyanuric acid, Alkalinity-M,	27 80 90	
3 - 150 mg/l O <sub>2</sub> (ø 16 mm)		0.02 - 4 mg/l Cl <sub>2</sub> / 6.5 - 8.4 pH	27.00.20	Calcium hardness, tablet reagen		
15 - 300 mg/l $O_2$ (ø 16 mm) available 20 - 1500 mg/l $O_2$ (ø 16 mm) 200 - 15000 mg/l $O_2$ (ø 16 mm)	soon!	powder reagents for chlorine 0.02 - 2.0 mg/l Cl <sub>2</sub> (ø 24 mm glass vial	o.c vder reagents for chlorine		$0.02 - 6.0 \text{ mg/l Cl}_2 / 0.1 - 10 \text{ mg/l Cl}_2*$ 0.05 - 13  mg/l Br; $6.5 - 8.4  pH0 - 160 \text{ mg/l cyanuric acid}; 5 - 200 \text{ mg/l CaCO}_3 (TA)$	
MD 100 Copper, tablet reagents 0.05 - 5.0 mg/l Cu	27 60 80	0.1 - 8.0 mg/l Cl <sub>2</sub> (ø 10 mm <b>multi vial-2</b> ) 6.5 - 8.4 pH		0 - 500 mg/l CaCO <sub>3</sub> (CaH)		
MD 100 Copper, powder reagents 0.05 - 5.0 mg/l Cu	27 60 85			MD 100 Boiler Water		
MD 100 Hardness, total, tablet reagents	27 61 90	3in1			27.62.20	
2 - 50 mg∕l CaCO₃		MD 100 Chlorine, pH, Cyanuric acid	<b>d</b> 27 80 10	MD 100 Aluminium, Chloride, Copper, DEHA, Hydra	27 62 30 zine,	
20 - 500 mg/l CaCO₃ (by dilution)  MD 100 Hazen, no reagents required 0 - 500 mg/l Pt-Co	27 61 60	tablet reagents <b>(OTZ)</b> 0.01 - 6.0 mg/l Cl <sub>2</sub> / 0,1 - 10 mg/l Cl <sub>2</sub> * 6.5 - 8.4 pH ; 0 - 160 mg/l cyanuric acid		Iron, Oxygen (dissolved), Phosphate, Polyacrylate, Silica (delivery without reagents)		
MD 100 Iron, tablet reagents 0.02 - 1.0 mg/l Fe	27 60 50	MD 100 Chlorine, pH, Cyanuric acid,	27 80 15			
MD 100 Iron TPTZ, powder reagents 0.02 - 1.8 mg/l Fe	27 60 55	liquid reagent for chlorine and pH <b>(OT</b> 0.02 - 4 mg/l Cl <sub>2</sub> / 6.5 - 8.4 pH 0 - 160 mg/l cyanuric acid	۷)	MD 100 Cooling Wate		
<b>MD 100 Iron</b> , powder reagents 0.02 - 3.0 mg/l Fe	27 60 56	MD 100 Chlorine, pH, Alkalinity-M tablet reagents (OTZ)	27 80 60	MD 100 Aluminium, Bromine, Chlorine, Chlorine HR, Chlorine	27 62 40 e dioxide,	
<b>MD 100 Fluoride</b> , without reagents 0.05 - 2.0 mg/l F	27 60 90	0.01 - 6.0 mg/l Cl <sub>2</sub> / 0,1 - 10 mg/l Cl <sub>2</sub> * 6.5 - 8.4 pH ; 5 - 200 mg/l CaCO <sub>3</sub> (TA)		Copper, Iron, Iron in Mo, Molyb Molybdate HR, Ozone, Polyacr		
MD 100 Manganese LR.	27 61 00			Sulphate, Triazoles, Zinc		

Chlorine, pH, Alkalinity-M (total)

0.01 - 6.0 mg/l  $\text{Cl}_2$ 5 - 200 mg/l  $\text{Cl}_2$  (ø 16 mm round vial) 0.02 - 11 mg/l  $\text{ClO}_2$ 

0.02 - 4 mg/l Cl<sub>2</sub> / 6.5 - 8.4 pH

5 - 200 mg/l CaCO<sub>3</sub> (TA)

Chlorine LR, Chlorine HR, Chlorine dioxide#, tablet reagents

liquid reagent for chlorine and pH (OTZ)

27 80 65

27 80 00

40 Molybdate HR, Ozone, Polyacrylate, Sulphate, Triazoles, Zinc (delivery without reagents)

- \* Delivery without reagents for measuring range 0.1 - 10 mg/l Cl<sub>2</sub>
- # Where chlorine and chlorine dioxide are present together, they may be determined quantitatively as a single figure.

MD 100 Manganese LR,

MD 100 Manganese LR,

MD 100 Manganese HR,

tablet reagents

0.2 - 4.0 mg/l Mn

powder reagents 0.01 - 0.7 mg/l Mn

powder reagents 0.1 - 18 mg/l Mn

27 61 00

27 61 05

27 61 06



# Scroll Memory (SM)

To avoid unnecessary scrolling for the required test method, the instrument memorizes the last method used before switching off the instrument. When the instrument is switched on again, the scroll list comes up with the last used test method first.

## **Delivery Content**

- Instrument in carrying case
- 4 micro batteries (AAA)
- 3 round vials (glass) with lid
- 1 stirring rod & 1 brush
- Tablet reagents and/or liquid reagents or VARIO Powder reagent
- Warranty information
- Certificate (COC)
- Instruction Manual

# Zero Setting (OTZ)

For certain versions of the instrument it is not necessary to zero the instrument each time. The zero setting is held in memory until the device is turned off. (One Time Zero - OTZ). The zero setting can be confirmed whenever it is required.

#### Manufacturers Test Certificate M

Besides the "Certificate of Compliance" which is supplied with the MD 100, manufacturers test certificates M are available at cost on request. Manufacturers test certificates M are individually supplied per instrument and per method.

The manufacturers test certificate M has to be ordered together with the new instrument and cannot be delivered at a later stage.

#### **N.I.S.T Traceability**

The device is factory pre-adjusted to international standards. The user can set the instrument in "user calibration mode" with standards traceable to N.I.S.T. adjust.

(N.I.S.T. = National Institute of Standards and Technology)

#### **Technical Data**

Optics	LEDs, interference filters (IF) and photo sensor in transparent sample chamber. Depending on the version, up to 3 different interference filters are used. Wavelength specifications of interference filters: 430 nm $\Delta \lambda = 5$ nm 530 nm $\Delta \lambda = 5$ nm 560 nm $\Delta \lambda = 5$ nm 580 nm $\Delta \lambda = 5$ nm 610 nm $\Delta \lambda = 6$ nm 660 nm $\Delta \lambda = 5$ nm	
Wavelength Accuracy	± 1 nm	
Photometric Accuracy <sup>4)</sup>	3 % FS (T = 20 °C – 25 °C)	
Photometric Resolution	0.01 A	
Power Supply	4 micro batteries (AAA), capacity approx. 17 hours or 5000 tests	
Auto - OFF	automatic switch-off	
Display	backlit LCD (on keypress)	
Storage	internal ring memory for 16 data sets	
Interfaces	infrared interface for test data transfer	
Additional feature	real time clock and date	
Calibration	factory calibration and user calibration. Reset to factory calibration possible	
Dimensions	155 x 75 x 35 mm (L x W x H)	
Weight	basic unit approx. 260 g	
Environmental conditions	temperature: 5–40°C rel. humidity: 30–90 % (non condensing)	

#### **CE-Conformity**

4) tested with standard solutions



#### Accessories

Accessories		
Item Set of 12 round vials with lid Height 48 mm, Ø 24 mm	<b>Code</b> 19 76	5 20
Set of 5 round vials with lid Height 48 mm, Ø 24 mm	19 76	5 29
Set of 10 round vials with lid Height 90 mm, Ø 16 mm	19 76	65
Adapter for round vials ø 16 mm	19 80	21 90
Set of 12 plastic vials (PC), with lid "Multi"-Type 2, Ø 10 mm	19 76	5 00
Vial stand for 6 round vials Ø 24 mm, acrylic glass	41 89	9 51
Vial stand for 10 vials (Ø 16 mm or □ 13,5 mm), acrylic glas	41 89 s	9 57
Mixing cylinder, 25 ml, with stopper required accessory for molybdenum L with MD 100 (276140)		26 50
Membrane filter set for use when preparing samples, 25 membrane filte 0,45 µm, 2 syringes 20 ml	36 6´ ers,	50
Cleaning cloth for vials	19 76	35
Set of 12 sealing rings for round vial ø 24 mm	19 76	5 26
4 micro batteries (AAA)	19 50	026
Measuring beaker, volume 100 ml	38 48	3 0 1
Plastic funnel with handle	47 10	07
Plastic stirring rod, 13 cm length	36 4 <i>′</i>	00
Plastic stirring rod, 13 cm length, (10 pc.)	36 4	1 20

Plastic stirring rod, 10 cm length

Infrared data transfer module IRiM

Plastic stirring rod, 10 cm length, (10 pc.) 36 41 30

36 41 09

21 40 50

# Primary chlorine standard kit

For checking each chlorine meter. This standard is easy to handle. US EPA Methode 334.0



ValidCheck Chlorine 1,5 mg/l Code: 48 10 55 10

#### Verification Standard Kit

The verification standard kit for the MD 100 is designed to assure the user of the accuracy and the reliability of the results related to the integrated wave lengths.

The kit contains one zero standard, 6 different vials for checking 6 different wave lengths and allows checking the complete range of MD 100 photometers.

The shelf life of the verification standard kit is two years from the date of production, provided that storage and use are in accordance with the instructions provided.

**Verification Standard Kit** 21 56 70

#### Reference Standard Kit for MD 100

The reference standards are designed to check the accuracy and the reliability of the results.

It is not possible to calibrate the photometer with the reference standards.

The shelf life of reference standards is two years from the date of production, provided that storage and use are in accordance with the instructions provided.

Kit Chlorine for instruments	27 56 50
with tablet / liquid reagent	
0.2* and 1.0* mg/l	
Kit Chlorine for instruments	27 56 55
with tablet / liquid reagent	
0.5* and 2.0* mg/l	
Kit Chlorine for instruments	27 56 56
with tablet / liquid reagent	
1.0* and 4.0* mg/l	
Kit Chlorine for instruments	27 56 60
with powder reagent	
0.2* and 1.0* mg/l	
Kit pH for instruments	27 56 70
with tablet / liquid reagent	
7.45* nH	

<sup>\*</sup> Approximate figure, actual figure specified in Certificate of Analysis







59